

DBH  
JPD

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 6/12/2015

---

Instrument Serial Number: 010521	As of 12-Jun-15
----------------------------------	-----------------

---

Start Date	End Date	Agency Name
11-Jun-15		Farmville PD
28-May-15	11-Jun-15	DFS Central Lab
17-Dec-14	28-May-15	Leesburg PD
02-Dec-14	17-Dec-14	DFS Central Lab
17-Jun-14	02-Dec-14	Patrick County Sheriff's Office
04-Jun-14	17-Jun-14	DFS Central Lab
04-Dec-08	04-Jun-14	Vinton PD
02-Jun-08	04-Dec-08	DFS Central Lab

DBH  
JPD

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010521 Worksheet Start Date 6/11/2015  
Location Farmville PD  
Address 116 N Main St Farmville VA 23901  
DFS Technician Donald Hall License No. 21250  
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 750 Reference Barometer (mm HG) 752  
Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

## Measurement Assurance Check

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.301
0.300	0.296	0.287	0.304	<b>Sample 2</b>	0.301
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.300
0.001		0.300	0.301		

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.099
0.100	0.099	0.096	0.102	<b>Sample 2</b>	0.099
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.) AG425201-31

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies

Mouthpieces

Certificates of Analysis

Operator Worksheet

Other:

50

50

Notes:

Instrument Serial Number

010521

Certification Date

06/12/15

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

6/12/15

Issuing Analyst

Jennie Ditty

Date

6/12/15

Intox EC/IR-II: Accuracy Check

DBH  
JPD

Department of Forensic Science

Serial Number: 010521 Test Number: 795

Test Date: 06/11/2015 Test Time: 13:13 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-12 Exp Date: 03/06/2017

Tank Pressure: 291 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	13:14
CHK	0.301	13:15
BLK	0.000	13:17
CHK	0.301	13:18
BLK	0.000	13:19
CHK	0.300	13:20

Test Status: Success

Calibration CRC: EB091AA5

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010521 Test Number: 796

Test Date: 06/11/2015 Test Time: 13:23 EDT

Dry Gas Target: 0.099

Lot Number: AG425201-31 Exp Date: 09/09/2016

Tank Pressure: 669 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	13:24
CHK	0.099	13:24
BLK	0.000	13:26
CHK	0.099	13:27
BLK	0.000	13:28
CHK	0.099	13:29

Test Status: Success

Calibration CRC: EB091AA5

DB1  
JRK

DBI  
JPL



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <u>INSTRUMENT, TEST,</u>	NAME OF COURT <u>DFS</u>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <u>HALL, DONALD, B</u>		AGENCY <u>DFS Central Lab</u>
DFS LICENSE NUMBER <u>21250</u>	LICENSE EXPIRES <u>10/01/2016</u>	DATE TEST CONDUCTED <u>06/11/2015</u>
TEST EQUIPMENT NUMBER <u>010521</u>		

RESULTS: TIME SAMPLE TAKEN 13:38 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:


I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE

FOR COURT



# IntoxNet MIS Report

Report Generated 12 Jun 2015 at 08:51

DBH  
JRD

## Test Results

Instrument Serial Number 010521

Test # 000797 Subject Test

Test Location 1 Department of  
Test Date 11 Jun 2015

Test Location 2 Forensic Science  
Test Time 13:32  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name HALL  
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

Effective Date 10/01/2014

License Number 21250

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 11 Jun 2015 End Time 13:39

Result Time 13:38

Result Date 11 Jun 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:32

Data Type BLK

Sample Value 0.000

Sample Time 13:33

Data Type CHK

Sample Value 0.098

Sample Time 13:34

Data Type BLK

Sample Value 0.000

Sample Time 13:35

Data Type SUBJ

Sample Value 0.000

Sample Time 13:36

Data Type BLK

Sample Value 0.000

Sample Time 13:37

Data Type SUBJ

Sample Value 0.000

Sample Time 13:38

Data Type BLK

Sample Value 0.000

Sample Time 13:39

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG425201-31

Standard Expiration Date 09/09/2016

Tank Pressure 659

Barometric Pressure 750 mmHg

Blow Sample Number 1 Blow Duration 3.39 sec

Blow Volume 1828 cc

End-of-Blow Time 13:36

Blow Sample Number 2 Blow Duration 4.33 sec

Blow Volume 2188 cc

End-of-Blow Time 13:38

Tamper Evident Stamp ea7fd848

Test Status Code 0

Test Status Success